§812.45

training and experience to investigate the device.

- (b) *Control of device*. A sponsor shall ship investigational devices only to qualified investigators participating in the investigation.
- (c) Obtaining agreements. A sponsor shall obtain from each participating investigator a signed agreement that includes:
- (1) The investigator's curriculum vitae.
- (2) Where applicable, a statement of the investigator's relevant experience, including the dates, location, extent, and type of experience.
- (3) If the investigator was involved in an investigation or other research that was terminated, an explanation of the circumstances that led to termination.
- (4) A statement of the investigator's commitment to:
- (i) Conduct the investigation in accordance with the agreement, the investigational plan, this part and other applicable FDA regulations, and conditions of approval imposed by the reviewing IRB or FDA;
- (ii) Supervise all testing of the device involving human subjects; and
- (iii) Ensure that the requirements for obtaining informed consent are met.
- (d) Selecting monitors. A sponsor shall select monitors qualified by training and experience to monitor the investigational study in accordance with this part and other applicable FDA regulations.

§812.45 Informing investigators.

A sponsor shall supply all investigators participating in the investigation with copies of the investigational plan and the report of prior investigations of the device.

§812.46 Monitoring investigations.

(a) Securing compliance. A sponsor who discovers that an investigator is not complying with the signed agreement, the investigational plan, the requirements of this part or other applicable FDA regulations, or any conditions of approval imposed by the reviewing IRB or FDA shall promptly either secure compliance, or discontinue shipments of the device to the investigator and terminate the investigator's participation in the investigation.

A sponsor shall also require such an investigator to dispose of or return the device, unless this action would jeopardize the rights, safety, or welfare of a subject.

(b) *Unanticipated adverse device effects.*(1) A sponsor shall immediately conduct an evaluation of any unanticipated adverse device effect.

- (2) A sponsor who determines that an unanticipated adverse device effect presents an unreasonable risk to subjects shall terminate all investigations or parts of investigations presenting that risk as soon as possible. Termination shall occur not later than 5 working days after the sponsor makes this determination and not later than 15 working days after the sponsor first received notice of the effect.
- (c) Resumption of terminated studies. If the device is a significant risk device, a sponsor may not resume a terminated investigation without IRB and FDA approval. If the device is not a significant risk device, a sponsor may not resume a terminated investigation without IRB approval and, if the investigation was terminated under paragraph (b)(2) of this section, FDA approval.

§ 812.47 Emergency research under § 50.24 of this chapter.

- (a) The sponsor shall monitor the progress of all investigations involving an exception from informed consent under §50.24 of this chapter. When the sponsor receives from the IRB information concerning the public disclosures under §50.24(a)(7)(ii) and (a)(7)(iii) of this chapter, the sponsor shall promptly submit to the IDE file and to Docket Number 95S-0158 in the Dockets Management Branch (HFA-305), Food and Drug Administration, 12420 Parklawn Dr., rm. 1-23, Rockville, MD 20857, copies of the information that was disclosed, identified by the IDE number.
- (b) The sponsor also shall monitor such investigations to determine when an IRB determines that it cannot approve the research because it does not meet the criteria in the exception in §50.24(a) of this chapter or because of other relevant ethical concerns. The sponsor promptly shall provide this information in writing to FDA investigators who are asked to participate in